

Irradiation certificate

For the attention of Mrs JEAN

ZI Les Chartinières

01220 Dagneux

CERTIFICATE N° 9502

In accordance with the Quality Assurance rules and with current legislation, IONISOS hereby certifies that it has performed Gamma radiation processing of the goods described below, to a minimum dose of:

25 kGy (kiloGrays)

Customer name: **Laboratoires MEDICAL**

Product name: **Medical equipment**

Customer order n°: **9903**

Number of packets: **800 packets**

Customer batch n°: **200B**

IONISOS reference n°: **99030**

IONISOS Batch(es)

and date of processing **C3200 of 15/02/00 800 packets**

At Dagneux on: **16/02/00**

Quality department stamp: